

Protocol Name: **ANKLE ROUTINE FOR LIG/NONSPECIFIC INJ.**

Body Part: **ANKLE**

Coil: **SMALL FLEX COIL**

Sequences	Seq Name	Image Plane	# Slices	TR	TE	Slice Thick	FOV	NSA	Matrix	Phase Dir
Axial Localizer	gre	Axial	5	7.8	3.69	6	350	1	256	A>P
3 Plane Localizer	gre	3 Plane	9	7.3	3.6	6	280	1	256	R>L
Sag T1 FS	tse	Sagittal	33	595	11	2.5	150	1	384	A>P
Sag STIR	tse	Sagittal	33	5670	65	2.5	150	1	320	A>P
Axial T1	tse	Axial Obl	50	610	15	2.5	150	1	384	R>L
Axial STIR	tse	Axial Obl	50	5180	55	2.5	150	2	320	R>L
Cor T1	tse	Coronal	45	655	15	3	150	1	384	R>L
Cor STIR	tse	Coronal	45	3950	67	3	150	1	320	R>L

DETAILS: The Oblique Axials need to cover from Metatarsal bases to just above the ankle joint at a 45 degree angle to the ankle joint, close to angle of calcaneus.

