

Protocol Name: **HINDFOOT FOR PLANTAR FASCIITIS**      Body Part: **FOOT**      Coil: **SMALL FLEX COIL**

Sequences	Seq Name	Image Plane	# Slices	TR	TE	Slice Thick	FOV	NSA	Matrix	Phase Dir
Axial Localizer	gre	Axial	5	7.8	3.69	6	350	1	256	A>P
3 Plane Localizer	gre	3 Plane	9	7.3	3.6	6	280	1	256	R>L
Sag T1 FS	tse	Sagittal	25	526	10	3	150	1	384	A>P
Sag PD FS	tse	Sagittal	25	2650	30	3	150	1	384	A>P
Axial Obl PD	tse	Axial Obl	40	650	11	3	150	1	384	A>P
Axial Obl T2 FS	tse	Axial Obl	40	5180	55	3	150	2	384	A>P
Cor T1	tse	Coronal	40	801	9.9	3.2	150	1	384	R>L
Cor T2 FS	tse	Coronal	40	3100	54	3.2	150	1	320	R>L