

Protocol Name: **PITUITARY**

Body Part: **BRAIN**

Coil: **HEAD**

Sequences	Sequence Name	# Slices	Image Plane	TR	TE	Slice Thickness	FOV	NSA	Matrix	Phase Dir
3 Plane Loc	gre	5	3 plane	8.6	4	7	250	2	256	A>P
Sag T1	gre	25	Sagittal	440	2.48	4	220	1	320	A>P
Cor T1 Thin	tse	15	Coronal	750	9	2	200	4	320	R>L
Sag T1 Thin	tse	15	Sagittal	550	9	2	200	2	320	A>P
Axial Diff	ep2d_diff	30	Axial	7827	100	4	220	4	192	A>P
Axial Flair	tse	30	Axial	9000	94	4	220	1	256	R>L
Gadolinium										
Sag T1 Thin Post	tse	15	Axial	550	9	2	200	2	320	A>P
Coronal T1 FS Thin Post	tse	15	Coronal	762	9	2	200	4	320	R>L
Axial T1 whole head Post	gre	25	Axial	250	2.48	4	220	2	320	R>L
Optional										
Cor T2 FS (post op)	tse		Coronal							

Details: Axial images should be done at 3mm / 0.5 skip
FOV should be 16 cm or less